



# SPEED™ IMPLANT

## TECHNIQUE OVERVIEW

Instructions for SPEED™ Memory Implants When Using DK-200, DU-200, DK-265, or DU-265 Drill Kits

# 1

Determine the correct Implant size by using the BME SPEED™ X-Ray Template or intraoperatively using the BME Continuous Compression Implant Sizing Guide (SG-1) to measure the fusion site.



# 2

Open the chosen Implant Kit and its corresponding Drill Kit. Connect the Drill Guide Tip, located in the Implant Kit, to the Drill Guide Handle located in the Drill Kit.



# 3

While ensuring full reduction, place the Drill Guide Assembly across the fusion site with both prongs touching bone. Drill the first hole using the Drill Bit provided in the Drill Kit until the positive stop is reached.



# 4

Insert a Pull Pin into the first hole and, while ensuring full reduction, repeat step 3 to create the second hole. **Optional:** Insert another Pull Pin into the second hole. The Drill Guide Assembly can be removed leaving the Pull Pins in place to mark the position of the drill holes. If desired, create a 1.0-1.5mm trough in line with the two drill holes so that the implant can be recessed.



# 5

Remove the Insertion Tool containing the SPEED Memory Implant from the Implant Package. Remove the Pull Pins from the pre-drilled holes and align the tips of the legs of the SPEED Memory Implant parallel with the drill holes.



# 6

Insert the SPEED Memory Implant as far as possible into the predrilled holes.

**Note:** To ensure proper implant placement, fluoroscopy may be used prior to releasing the implant.



# 7

Press the distal third of the central button with thumb while simultaneously twisting the insertion tool until disengagement has been verified.

Additional implant disengagement options:

- Press button with thumb to release.
- Twist Insertion Tool in either direction to release.



# 8

Align the supplied Tamp with the bridge of the implant and use as needed to completely seat the implant.



# 9

Repeat steps 1-8 for each additional implant used.

**Tip:** If implants are placed at 90 degrees to each other, stagger them to ensure unobstructed insertion. If a second implant is placed with the bridge in close proximity to another, insert the implant with the open side facing the first Implant. This allows unobstructed release of the Implant from the Insertion Tool.

Limited Warranty and Disclaimer: DePuy Synthes Products are sold with a limited warranty to the original purchaser against defects in workmanship and materials. Any other express or implied warranties, including warranties of merchantability or fitness, are hereby disclaimed.

CAUTION: Federal Law restricts these devices to sale by or on the order of a physician. Some devices listed in this surgical technique overview may not have been licensed in accordance with Canadian law and may not be for sale in Canada.

Please contact your sales consultant for items approved for sale in Canada.

Not all Products may currently be available in all markets.

The above steps are an overview of the surgical technique. Complete information regarding indications, contra-indications, warnings, care and caution can be found in the Instructions For Use.

**REMOVAL:** 1. Expose the site and the bridge of the implant. 2. Using forceps, grasp the center of the implant and remove. If the implant is recessed, then use an elevator to lift the implant bridge and then use forceps to remove the implant. If solidly connected, straight implants can be removed by cutting the center of the bridge of the implant and removing the remnants with an elevator.



**Manufactured or Distributed By:**

**Synthes USA, LLC**  
1101 Synthes Avenue  
Monument, CO 80132

**BioMedical Enterprises, Inc.**  
14785 Omicron Dr., Suite 205  
San Antonio, TX 78245

*NOTE: For recognized manufacturer, refer to the product label.*

[www.depuyssynthes.com](http://www.depuyssynthes.com)

© DePuy Synthes 2016. All rights reserved.  
DSUS/TRM/0716/0955(1) 10/16

**TO ORDER THIS PRODUCT,  
PLEASE CONTACT:**



**BioMedical Enterprises**

(800) 880-6528 Toll Free  
(866) 913-3977 Fax  
(210) 677-0354 Corporate  
sales@bme-tx.com Email

[www.bme-tx.com](http://www.bme-tx.com)